510(k) Summary

as required by 807.92

K030276

1. Company Identification

Totoku Electric Co., Ltd.

300 Oya, Ueda-shi, Nagano-ken, 386-0192, JAPAN

Tel: 011-81-268-34-5484

Fax: 011-82-268-34-5565

2. Official Correspondent

Mikio Hasegawa (Mr.)

General Manager

Product Development Dept.

3. Date of Submission

Jan. 24, 2003

4. Device Trade Name

Flat Panel Displays, CDL Series

5. Common Name

Monitor, display, workstation, and others

6. Classification

Medical displays are classified as Class I or II per 21 CFR 890.2050

7. Predicate Device

Totoku ME311L 3 Mega Pixel Diagnostic Display, manufactured by Totoku Electric Co., Ltd. (K012099). Comparison of the principle characteristics of the device which is pertinent to clinical performance is shown in Appendix 1.

8. Description of Device

CDL Series Medical Displays are displays for medical use.

9. Intended Use

CDL Series Medical Displays are intended for use in viewing digital medical images.

10. Explanation of CDL Series

CDL Series consists of color LCD displays listed below.

Model No. CDL1811A Model No. CDL1813A

Comparison of specifications are shown in Appendix 2.

11. Compliance

All CDL Series listed above comply with the following standards.

Medical Safety: UL2601-1, CSA No. 601-1, MDD/CE (EN60601-1, IEC60601-1)

EMC: MDD/CE (EN60601-1-2), IEC60601-1-2, and FCC-B

Appendix 1 Specification Comparison Chart with Predicate Device

| Item | ME311L | CDL1811A |
|---|--|--|
| 510(k) Number | K012099 | Not known |
| Display area | Horizontal:423.9mm, Vertical:318.0mm | Horizontal: 359.0mm, Vertical: 287.2mm |
| Input signal | GVIF video signal 10214-1210VE (3M:MDR14P) | D-SUB (analog), DVI-D (digital) |
| Maximum display pixels | Portrait: 1536 dots X 2048 line Landscape: 2048 dots X 1536 line | 1280 X 1024 dots |
| Scanning frequency | Horizontal:93KHz, Vertical:60Hz | Horizontal: 31K - 80KHz, Vertical: 55 - 85Hz |
| Maximum image clock | 65MHz | 135MHz |
| Maxmum brightness | 600cd/m2 | 240cd/m2 |
| Brightness calibration | Software(option) Photosensor(option item)-DTP92(X-Lite) | The second secon |
| Serial communication connector | D-sub 9P x 2 | 15P Mini D-SUB, 24P DVI-D, 4P DC input terminal 9p Mini D-SUB serial (RS232C) |
| | Medical safety:UL2601-1,CSA No.601-1 EN60601-1 MDD/CE:(EN60601-1, EN60601-1-2) | Medical safety: UL2601-1, CSA CSA C22.2No.601.1 (EN60601-1), FCC Class B, DOC-B, BSMI |
| Dimensions and weight (incl. Tilt and swivel) | Net, 486x480x250mm(W x H x D) (landscape) 11kg 380x533x250mm(W x H x D) (portrait) Packed, 733x642x363mm(W x H x D) 17kg | Net: 432x353x68.6mm(W x H x D) (landscape) 6.9kg |
| Power supply | 100-240V AC, 50/60Hz | Packed: 485x600x280mm(W x H x D) 12kg 100-250V AC, 50/60Hz |

Appendix 2 Specification Comparison Chart of the Applied Models

| Item | ODL1811A | CDL1813A |
|--------------------------------|--|--|
| 510(k) Number | Not known | Not known |
| Display area | Horizontal: 359.0mm, Vertical: 287.2mm | Horizontal: 359.0mm, Vertical: 287.2mm |
| Input signal | D-SUB (analog), DVI-D (digital) | D-SUB (analog), DVI-D (digital) |
| Maximum display pixels | 1280 X 1024 dots | 1280 X 1024 dots |
| Scanning frequency | Horizontal: 31K - 80KHz, Vertical: 55 - 85Hz | Horizontal: 31K - 80KHz, Vertical: 55 - 85Hz |
| Maximum image clock | 135MHz | 135MHz |
| Maxmum brightness | 240cd/m2 | 240cd/m2 |
| Brightness calibration | | 2 1000/ 1112 |
| Serial communication connector | 15P Mini D-SUB, 24P DVI-D, 4P DC input terminal 9p Mini D-SUB serial (RS232C) | 15P Mini D-SUB, 24P DVI-D, 4P DC input terminal 9p Mini D-SUB serial (RS232C) |
| Agency standards | | Medical safety: UL2601-1, GSA GSA C22.2No.601.1 (EN60601-1), FCC Class B, DOC-B, BSMI |
| Dimensions and weight | Net: 432x353x68.6mm(W x H x D) (landscape) 6.9kg | Net: 432x353x68.6mm(W x H x D) (landscape) 6.9kg |
| | Packed: 485x600x280mm(W x H x D) 12kg | Packed: 485x600x280mm(W x H x D) 12kg |
| Power supply | 100-250V AC, 50/60Hz | 100-250V AC, 50/60Hz |



APR 2 5 2003

Food and Drug Administration 9200 Corporate Boulevard Rockville MD 20850

Mr. Mikio Hasegawa Re: K030276
General Manager Trade/Dev
Totoku Electric Co., Ltd. Regulation
300 Oya, Ueda-shi
Nagano 286-0192 Regulator
JAPAN Resultator
Product C

Trade/Device Name: Flat Panel Display, CDL Series

Regulation Number: 21 CFR 892.2050
Regulation Name: Picture archiving and communications system

Regulatory Class: II Product Code: 90 LLZ Dated: January 24, 2003 Received: January 27, 2003

Dear Mr. Hasegawa:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at one of the following numbers, based on the regulation number at the top of the letter:

| 8xx.1xxx | (301) 594-4591 |
|----------------------------------|----------------|
| 876.2xxx, 3xxx, 4xxx, 5xxx | (301) 594-4616 |
| 884.2xxx, 3xxx, 4xxx, 5xxx, 6xxx | (301) 594-4616 |
| 892.2xxx, 3xxx, 4xxx, 5xxx | (301) 594-4654 |
| Other | (301) 594-4692 |

Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97) you may obtain. Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address http://www.fda.gov/cdrh/dsma/dsmamain.html.

Sincerely yours,

Mancy C. Brogdon
Nancy C. Brogdon

Director, Division of Reproductive, Abdominal and Radiological Devices

Office of Device Evaluation

Center for Devices and Radiological Health

Enclosure

| Device Name: | Flat Panel Displays, CDL Seri | es | |
|--------------------|---|--|---|
| Indications for Us | e: | | |
| CDL Series Medi | cal Displays are intended for use | in viewing digital medical images. | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| (PLEASE DO NOT WI | RITE BELOW THIS LINE-CO | NTINUE ON ANOTHER PAGE IF NECESSARY |) |
| | Concurrence of CDRH, Office of | of Device Evaluation | |
| | (Division Sign Division of Re and Radiologic 510(k) Number | Off) Productive, Abdominal, cal Devices K030276 | |
| Prescription Use | | OR Over-The-Counter Use | |
| | | (Optional Format 1-2-96) | |

Not known

510(k) Number (If known):

K030276